



स्वास्थ्य मन्त्रालय

बोलपत्र/दरभाउपत्र सम्बन्धी सूचना

दोश्रो पटक प्रकाशित मिति २०७४।०५।२८

आ.व.०७४।०७५ मा यस अस्पतालको लागि आवश्यक Bed Side Patient Monitor बोलपत्रको माध्यम वाट खरीद गर्नुपर्ने भएकोले योग्य इजाजत प्राप्त बोलपत्रदाता वाट निम्नशर्तहरूको अधिनमा रही रित पूर्वकको बोलपत्र भरतपुर अस्पतालको Web site www.bharatpurhospital.gov.np डाउनलोड गरि खरीद गर्न सकिनेछ ।

शर्तहरू:

१. आ.व. २०७३/०७४ का लागि नविकरण भएको इजाजतपत्रदर्ता र मु.अ.कर/प्यान नं. सहितको प्रमाण पत्रको प्रमाणित प्रतिलिपी संलग्न राखिनिवेदन दिई भरतपुर अस्पतालको Web site www.bharatpurhospital.gov.np मा गई TOR download गर्न सकिनेछ । बोलपत्र खरीद गर्न सूचना प्रकाशित भएको मितिले तपशिल अनुसारको रकम, उल्लेखित मिति र समयमा यस अस्पतालको रा.वा.बैंक भरतपुरमा रहेको चल्तीखाता नं.१३१००००९६७०१ मा जम्मा गरी खरीद गर्न सकिनेछ ।
२. बोलपत्र पेश गर्दा जमानत वापत तपशिलमा उल्लेखित रकम को.ले.नि.का.चितवनको रा.वा.बैंक, भरतपुर शाखामा रहेको ख ३ धरौटी १३१०२०३०००००० खातामा रकमजम्मा गरेको सक्कल बैंक भौचर वा कम्तिमा १२० दिन म्याद भएको यस अस्पतालको नाममा जारी भएको विड वण्ड पेश गर्नुपर्नेछ । बोलपत्र स्वीकृत भए पछि नियमानुसार आवश्यक यस कार्यालयको नाममा जारी भएको कम्तिमा एक वर्ष म्याद भएको बैंक परफरमेन्स वण्ड पेश गर्नुपर्नेछ ।
३. खरीद भएको बोलपत्र/दरभाउपत्र तपशिलमा उल्लेखित मिति र समयमा भरतपुर अस्पतालको Web site www.bharatpurhospital.gov.np मा गई बैंक भौचर, विडवण्ड, BOQ समेतका कागजात Scan गरी Upload गरी सक्नुपर्नेछ । Upload भएका बोलपत्र तपशिलमा उल्लेखित मिति र समयमा यस अस्पतालमा बोलपत्रदाता वा निजको प्रतिनिधिको रोहवरमा खोलिने छ । प्रतिनिधि उपस्थित नभएपनि बोलपत्र खोल्न बाधा पर्नेछैन । बोलपत्र खरीदगर्ने वा दर्ता गर्ने अन्तिम दिन विदा पर्न गएमा सोको भोलिपल्ट क्रमश उक्त कार्य हुनेछ ।
४. बोलपत्र दाताले कबोल गरेको प्रत्येक आइटमको दररेटलाई (कर बाहेक) अंक र अक्षरमा स्पष्टसँग उल्लेख गर्नुपर्नेछ । अंक र अक्षरमा उल्लेखित दररेटमा भिन्नता भएमा अक्षरलाई मान्यतादिइने छ ।
५. बोलपत्र दाताले कबोल गरेको बोलपत्रमा तोकिएको परिमाण मापदण्ड अनुसार र समयमा सामान सप्लाई गर्न नसकेमा बजारबाट सिधै खरीद गरिनेछ । त्यसरी खरीद गर्दा बोलपत्रमा उल्लेख गरिएको मुल्यभन्दा बढी पर्न गएमा बोलपत्र दाताको धरौटी रकम वा भुक्तानी गर्न बाकि रकमबाट असुलउपर गरिनेछ ।
६. बोलपत्रमा उल्लेखित सामानहरू पटक पटक वा एकमुष्ट रुपमा खरीद गरिने छ । खरीद गरिने सामानको परिमाण आवश्यकता अनुसार हुनेछ ।
७. कुनै बोलपत्र स्वीकृत गर्ने वा नगर्ने वा आइटम वाइज रुपमा स्वीकृत गर्ने सम्पूर्ण अधिकार यस अस्पतालमा सुरक्षित रहनेछ ।
८. संभौता गरी सामान सप्लाई नगर्ने फर्मलाई नियमानुसार कालोसूचीमा दर्ताको लागि सम्बन्धित निकायमा लेखि पठाइनेछ ।
९. थप जानकारीको लागि यस अस्पतालको जिन्सी शाखामा सर्म्पर्क राख्न सकिनेछ ।

Medical Superintendent



- १०) शर्त रहेको, रीत नपुगेको तथा म्याद नाघी आएको बोलपत्र उपर कुनै कारवाही हनेछैन । एक फर्म वा सस्थाको नाममा खरिद गरिएको बोलपत्र अर्को फर्मको नामवाट जम्मा गर्न पाईने छैन ।
- ११) बोलपत्रदाताले आफू कालोसूचीमा नपरेको भनी स्वयं घोषणा पेश गर्नुपर्नेछ ।
- १२) औषधीजन्य सामाग्रीहरुको हकमा मूल्य अभिवृद्धि कर/प्यान र सर्जिकलजन्य सामानहरुको हकमा मु.अ.करमा दर्ता भएको हुनुपर्नेछ ।
- १३) यसमा नपरेका अन्य कुराहरु सार्वजनिक खरीद ऐन २०६३ र सार्वजनिक खरीद नियमावली २०६४ अनुसार हुनेछ ।

सि.नं.	ठे.नं.	विवरण	धरौटी रकम	बोलपत्र दस्तुर	बोलपत्र दाखिला गर्ने अन्तिम मिति	बोलपत्र खोले मिति	बोलपत्र दस्तुर दाखिला गर्ने बैक तथा खाता नं.	कैफियत
१	२११०७४१०७५	Bed Side Patient Monitor	१०००००।	३०००।००	१५ औं दिनको ५१०० वजे	१६ औं दिनको २।०० वजे	रा.वा.बैक भरतपुर, च.हि.नं ९६७	

Medical Superintendent



Bedside Patient Monitor

S.N.	Purchaser's Specifications	Bidder's Offer
	Bedside Patient Monitor	
	Manufacturer	
	Brand	
	Type / Model	
	Country of Origin	
1	Description of Function	
1.1	For monitoring vital signs of all patient categories, at bedside or during transportation.	
2	Operational Requirements	
2.1	It shall operate on AC power supply as well as built-in battery.	
3	System Configuration	
3.1	Patient Monitor, portable with complete accessories.	
4	Technical Specifications	
4.1	Monitor should work in paediatric to adult.	
4.2	Mounting solution / mounting on any pendant for monitors	
4.3	Parameters monitored: ECG, Respiration Rate, SpO2, Noninvasive blood pressure (NIBP) and Temperature measurements.	
4.4	24 hours real time ECG waveform storage and replay.	
4.5	12/24/96-hour trend graph and data storage.	
4.6	NIBP oscillometric step deflation, manual/automatic, initial inflation pressure user selectable	
4.7	High resolution TFT color display with at least 12" display or more.	
4.8	Enlarged vital sign values to facilitate viewing from distance	
4.9	Sweep adjustable to 12.5, 25 or 50mm/second.	
4.10	Sensitivity (amplitude) of all signals user adjustable.	
4.11	Standardising marker, 1mV.	
4.12	User pre-set of high/low alarms on all monitored parameters.	
4.13	Audio-visual alarm in case measurements are outside pre-set range.	
4.14	Silencing feature for audio alarms.	
4.15	Safe to work with defibrillator, electric knife and pacemaker	
4.16	Battery back up for at least 2 hours	
5	Accessories, spares and consumables	
5.1	Accessories:	
	1 x Mounting bracket for fixation to standard bed/wall rail and mobile pole/stand NIBP accessories: <ul style="list-style-type: none"> • 1 x NIBP hose • Blood pressure cuff (1 x large adult, 1 x medium adult, 1 x paediatric size) ECG accessories <ul style="list-style-type: none"> • 1 set 3 lead cables • 1 set 5 lead cables • 1 x Electrode gel, bottle 350ml 	

Medical Superintendent



	Temperature accessories <ul style="list-style-type: none"> • 1x Skin temperature probe (including connection cable) Pulse oximetry (SpO2) sensors with cable and plug <ul style="list-style-type: none"> • 2 x Adult size, reusable clip-on type/ Silicon type • 2 x Infant size, reusable clip-on type/Silicon type 	
S.N.	Purchaser's Specifications	
5.2	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).	
6	Operating Environment	
6.1	The system offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The condition include Power Supply, Climate, Temperature Humidity, etc	
6.2	Power supply: 220 – 240 V AC, 50Hz fitted with appropriate plug. The power cable must be atleast 3m in length.	
7	Standards and Safety Requirements	
7.1	USFDA approved product certificate is mandatory.	
7.2	Shall meet IEC-60601-1-2:2001 General Requirements of Safety for Electromagnetic Compatibility.	
7.3	Safety requirements as per IEC 60601-2-27:1994—Medical electrical equipment—Part 2: Particular requirements for the safety of electrocardiographic monitoring equipment.	
8	User Training	
8.1	Must provide user training (including how to use and maintain the equipment).	
9	Warranty	
9.1	Comprehensive warranty for 2 years after acceptance.	
10	Maintenance Service During Warranty Period	
10.1	During the warranty period supplier must ensure planned preventive maintenance (PPM) along with corrective/breakdown maintenance whenever required.	
11	Installation and Commissioning	
11.1	The bidder must arrange for the equipment to be installed and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail	
12	Documentation	
12.1	User (Operating) manual in English.	
12.2	Service (Technical / Maintenance) manual in English	
12.3	List of important spare parts and accessories with their part number and costing.	
12.4	Certificate of calibration and inspection from factory.	

Medical Superintendent



स्वास्थ्य मन्त्रालय

भरतपुर अस्पताल

भरतपुर, चितवन

Bedside Patient Monitor को बोलपत्र

S.N	Generic Name	Unit	Qty	दररेट		कैफियत
				अंकमा	अक्षरमा	
1	Bedside Patient Monitor	PC	20			
Total						
VAT 13%						
Grand Total						
	Name of Bidder:					
	Signature:					
	Date					
	Office Seal					

Medical Superintendent