



नेपाल सरकार
स्वास्थ्य तथा जनसंख्या मन्त्रालय
भरतपुर अस्पताल, चितवन
वोलपत्र सम्बन्धी सूचना
(दोश्रो पटक प्रकाशित मिति २०७७/०१/२३)

यस अस्पताल लाई चालु आ.व. ०७६/०७७ को लागि आवश्यक पर्ने RVG Sensor System बोलपत्रको माध्यम बाट खरिद गर्नुपर्ने भएकोले मु.अ.क.मा दर्ता भएका इच्छुक ईजाजत प्राप्त सप्लायर्स, विक्रेता तथा व्यवसायीहरूले रित्तपूर्वकको बोलपत्र तपशिलको शर्तहरू पालना हुने गरी यस अस्पतालको Website- www.bharatpurhospital.gov.np को eProcurement System मार्फत पेश गर्नु हुन यो सूचना प्रकाशित गरिएको छ ।

शर्तहरू:

१. आ.व. २०७६/०७७ का लागि नविकरण भएको इजाजतपत्रदर्ता र मु.अ.कर दर्ता सहितको स्थायी लेखा नं. प्रमाण पत्र, आ.व.०७५/०७६ को करचुक्ता को प्रमाणपत्र पेश गर्नु पर्ने छ ।
२. बोलपत्र खरिद तथा पेश गर्नको लागि तपशिल अनुसारको बोलपत्र दस्तुर रकम, उल्लेखित मिति र समयमा यस अस्पतालको रा.बा.बैंक भरतपुरमा रहेको चल्लीखाता नं.१३१००००९६७०१ मा जम्मा गरेको बैंक भौचर पेश गर्नु पर्ने छ ।
३. बोलपत्र पेश गर्दा जमानत वापत तपशिलमा उल्लेखित रकम को.ले.नि.का.चितवनको रा.बा.बैंक, भरतपुर शाखामा रहेको ख ३ धरौटी १३१०२०३०००००० खातामा (जम्मा गर्न पठाउने कार्यालयको नाम-भरतपुर अस्पताल,चितवन, कोड नं.३७०००३१०२ लेखि) रकम जम्मा गरेको सककल बैंक भौचर वा कम्तिमा १२० दिन म्याद भएको यस अस्पतालको नाममा जारी भएको विड वण्ड पेश गर्नुपर्नेछ ।
४. बोलपत्र तपशिलमा उल्लेखित मिति र समयमा यस अस्पतालको Website- www.bharatpurhospital.gov.np को eProcurement System मार्फत अनिवार्य रुपमा पेश गर्नु पर्ने छ
५. Upload भएका बोलपत्रहरू तपशिलमा उल्लेखित मिति र समयमा यस अस्पतालमा बोलपत्रदाता वा निजको प्रतिनिधिको रोहवरमा खोलिने छ । प्रतिनिधि उपस्थित नभएपनि बोलपत्र खोल्न बाधा पर्ने छैन । बोलपत्र खरीदगर्ने वा दाखिला गर्ने अन्तिम दिन विदा पर्न गएमा सोको भोलिपल्ट क्रमश उक्त कार्य हुनेछ ।
६. बोलपत्र दाताले कबोल गरेको प्रत्येक आइटमको दररेटलाई (मु.अ.कर बाहेक) अंक र अक्षरमा स्पष्टसँग उल्लेख गर्नुपर्नेछ । अंक र अक्षरमा उल्लेखित दररेटमा भिन्नता भएमा अक्षरलाई मान्यतादिइने छ ।
७. बोलपत्र दाताले कबोल गरेको बोलपत्रमा तोकिएको परिमाण मापदण्ड अनुसार र समयमा सामान सप्लाइ गर्न नसकेमा बजारबाट सिधै खरीद गरिनेछ । त्यसरी खरीद गर्दा बोलपत्रमा उल्लेख गरिएको मुल्यभन्दा बढी पर्न गएमा बोलपत्र दाताको धरौटी रकम वा भुक्तानी गर्न बाकि रकमबाट असुल उपर गरिनेछ ।
८. सप्लाइ गरिने सम्पूर्ण सामानहरू स्पेसिफिकेशन वमोजिमको उच्च गुणस्तरको हुनपर्नेछ ।


Medical Superintendent



१. कुनै बोलपत्र स्वीकृत गर्ने वा नगर्ने वा आंशिक रूपमा स्वीकृत गर्ने सम्पूर्ण अधिकार यस अस्पतालमा सुरक्षित रहनेछ ।
१०. सप्लाई गरिएका सामानहरू निरिक्षण एवं परिक्षण गर्दा बोलपत्र पेश गर्दा वखतको स्पेसिफिकेशन तथा गुणस्तर बमोजिमको नभए सम्मौता रद्द गरी कारवाहि गरिनेछ ।
११. संभौता गरी सामान सप्लाई नगर्ने फर्मलाई धरौटी जफत गरी नियमानुसार कालोसूचीमा दर्ताको लागि सम्बन्धित निकायमा लेखि पठाइनेछ ।
१२. शर्त रहेको, रीत नपुगेको तथा म्याद नाघी आएको बोलपत्र उपर कुनै कारवाही हुनेछैन । एक फर्म वा सस्थाको नाममा खरिद गरिएको बोलपत्र अर्को फर्मको नामवाट जम्मा गर्न पाइने छैन ।
१३. बोलपत्रदाताले खरिद कारवाहिमा भाग लिन प्रचलित कानून अनुसार अयोग्य नभएको, प्रस्तावित खरिद कारवाहिमा आफ्नो स्वार्थ नवाभिएको वा व्यवसाय सम्बन्धि कसुरमा आफुले सजाय नपाएको भनि स्वयं घोषणा पेश गर्नुपर्नेछ । त्यस्तो स्वयं घोषणा सूचना प्रकाशित भएको मिति पछिको हुनु पर्नेछ ।
१४. बोलपत्रदाताले स्पेसिफिकेशन राम्रो सँग अध्ययन गरि सोहि बमोजिम वारेन्टि, ग्यारेन्टि पश्चात वार्षिक मर्मत संभार शुल्क समेतलाई मध्यनजर गरी दररेट पेश गर्नु पर्नेछ ।
१५. स्वीकृत बोलपत्रदाताले मेशिनरी उपकरण जडान गर्दा लाग्ने थप एसोसेरिज सामानहरू स्वयं आफैले व्यवस्था गरी जडान गरी संचालनमा ल्याउनुपर्ने छ ।
१६. यसमा नपरेका अन्य कुराहरू सार्वजनिक खरीद ऐन २०६३ र सार्वजनिक खरीद नियमावली २०६४ अनुसार हुनेछ ।
१७. थप जानकारीको लागि यस अस्पतालको जिन्सी शाखा साथै लेखामा सर्म्पक राख्न सकिनेछ ।

सि.नं.	ठे.नं.	विवरण	जमानत रकम	बोलपत्र दस्तुर	बोलपत्र खरिद गर्ने अन्तिम मिति	बोलपत्र दाखिला गर्ने अन्तिम मिति	बोलपत्र खोल्ने मिति	कैफियत
१	३०,०७६।०७७	RVG Sensor System	१२०००.००	१०००.००	१५ औँ दिनको ४१०० वजे	१६ औँ दिनको १२।०० वजे	१६ औँ दिनको १।०० वजे	पुन प्रकाशन


Medical Superintendent



नेपाल सरकार

स्वास्थ्य तथा जनसंख्या मन्त्रालय

भरतपुर अस्पताल, चितवन

आ.व.०७६।०७७ को लागि आवश्यक पर्ने RVG Sensor System को (ठे.नं. ३०,०७६।०७७) को वोलपत्र फाराम

सि.नं.	सामानको नाम	ईकाइ	आवश्यक संख्या	प्रति ईकाइ दर रु अंकमा	प्रति ईकाइ दर रु अक्षरमा	जम्मा रकम रु अंकमा	कैफियत
१	RVG Sensor System	Set	१				
जम्मा							
मु.अ.क. थप १३ प्रतिशत							
कुल जम्मा मु.अ.क समेत							
वोलपत्र पेश गर्ने फर्मको नाम, ठेगाना-							
आधिकारिक दस्तखत-							
फर्म(कम्पनी) को आधिकारिक छाप-							
वोलपत्र पेश गर्दाको मिति-							

Medical Superintendent



RVG System

S.N.	Purchaser's Specifications	Bidder's Compliance Sheet			
		Yes	No	Page No. in Catalogue	Remarks
1	RVG Sensor System				
	Manufacturer				
	Brand				
	Type / Model				
	Country of Origin				
2	Description of Function				
2.1	RVG is a Digital dental imaging system, which allow quick or immediate viewing of images without using dental x-ray film, consist of an intraoral sensor or imaging plate, an x-ray system, computer hardware and software for image processing, and a hard-copy printer.				
3	Operational Requirements				
3.1	RVG Sensor and the Computer system along with imaging software are required. X-Ray generator is not to be quoted.				
4	System Configuration				
4.1	RVG Sensor System complete unit.				
5	Technical Specifications				
5.1	Must be based on CMOS / APS.				
5.2	10-30 lp/mm true image resolution.				
5.3	Exclusive sensor with complete software package including optical fibre technology.				
5.4	Available in three (Small, Medium & Large) sizes to help meet the unique imaging needs of practice and patients.				
5.5	Plastic pack design to allow easy periapical and bitewing radiograph.				
5.6	Thickness of the sensor must be 2-5 mm.				
5.7	Sensor life must be more than 350000 Exposures.				
5.8	Sensor active area must range from 450-1000 Square mm for different sizes of sensors.				

Medical Superintendent



5.9	RVG System Software.				
5.10	RVG/System software shall have: <ul style="list-style-type: none"> • Must be licensed. • Must have facility for RVG as well as intra oral Camera. • Must have automatic acquisition and save facility. • Must have sharpening, cleaning and improving features. • Must have search facility by patient ID name or any other criteria. • Must be Internet compatible. • Must be capable of generating Reports. • Must be capable of avoiding accidental deletion 				
5.11	Computer Hardware				
	Computer Hardware: <ul style="list-style-type: none"> • Must be Intel Core i5 Pentium PC with 500 GB Hard disk, Min.4GB RAM, DVD writer, 17 inches LCD/TFT Monitor. • Shall have compatible space-saving colour dye-sublimation printer , 400dpi resolution, A6-sized images in less than 25 seconds or equivalent. 				
7	Accessories, spares and consumables				
7.1	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer.				
7.2	Accessories: <ul style="list-style-type: none"> • Lead apron, thyroid collar and gonadal sheath-2 each • Computer with colour photo printer as specified • Suitable UPS with maintenance free batteries for minimum one-hour back-up must be supplied with the system. 				
8	Operating Environment				
8.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power Supply, Climate, Temperature, Humidity, etc.				
8.2	Power supply: 220 – 240 VAC, 50Hz fitted with appropriate plug. The power cable must be at least 3 metre in length.				


Medical Superintendent



8.3	UPS of suitable rating with voltage regulation, spike protection and maintenance free batteries for 30 minutes back up				
10	Standards and Safety Requirements				
10.1	Must submit ISO13485:2003/AC:2007 for Medical Devices AND				
10.4	CE (93/42 EEC Directives) or USFDA approved product certificate.				
	Shall meet IEC-60601-1-2:2001 General Requirements of Safety for electromagnetic Compatibility or must comply with 89/366/EEC; EMC directive.				
11	User Training				
11.1	Must provide user training (including how to use and maintain the equipment).				
12	Warranty				
12.1	Comprehensive warranty for 2 years.				
12.2	Maintenance Service During Warranty Period				
12.3	During the warranty period supplier must ensure corrective/breakdown maintenance whenever required. (written document)				
13	Installation and Commissioning				
13.1	The bidder must arrange for the equipment to be installed and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.				
14	Documentation				
14.1	User (Operating) manual in English.				
14.2	Service (Technical / Maintenance) manual in English.				
14.3	Certificate of calibration and inspection from factory.				
14.4	List of important spare parts and accessories with their part number and costing.				

Bidder must completely fill the Technical Specification Form (TSF). Only Yes/no/all complies should not be written. Page number in the catalogue of all the required parameters must be clearly mentioned and highlighted. Failure in doing so may lead to rejection of bid from technical committee.


 Medical Superintendent