



#### स्वास्थ्यमन्त्रालय

## बोलपत्र / दरभाउपत्र सम्बन्धी सूचना

### प्रथम पटक प्रकाशितमिति २०७३।१२।२८

चालु आ.ब.०७३१०७४ मायस अस्पताललाई आवश्यक पर्ने तपिशल वमोजिमका मेशिनरी औजार तथा उपकरणहरू बोलपत्र/दरभाउपत्रको माध्यमद्धारा आपूर्ति गर्नुपर्ने भएकोले तपिशल वमोजिमको शर्तमा रहीबोलपत्र/दरभाउपत्र पेश गर्नुहुन सबैको जानकारीको लागि यो सुचनाप्रकाशितगरिएको छ ।

#### शर्तहरुः

- १. बोलपत्र/दरभाउपत्र खरिद गर्न मु.अ.करमा दर्ता भएका, आ.ब.०७३१०७४ को लागि निवकरण भएको र सो कार्य सम्बन्धी इजाज तप्राप्त सप्लायर्स/विक्रेता तथा व्यवसायीहरुले रितपूर्वकको बोलपत्र/दरभाउपत्र सार्वजिनक खरिद अनुगमन कार्यालयको http//www.bolpatra.gov.np मागई TOR downloadगर्न सिकनेछ ।TOR खरीदगर्न सूचना प्रकाशित भएको मितिले तपिशल अनुसारको रकम, उल्लेखित मिति र समयमा यस अस्पतालको रा.बा.बैंक भरतपुरमा रहेको चल्ती खाता नं.१३१००००९६७०१मा नगद जम्मा गरी खरीद गर्न सिकनेछ ।
- २. बोलपत्र पेश गर्दा जमानत वापत तपिशलमा उल्लेखित रकमको.ले.नि.का.चितवनको रा.बा.बैंक, भरतपुर शाखामा रहेको ख ३ धरौटी १३१०२०३००००० खातामा रकम जम्मा गरेको सक्कल बैंक भौचर वाकम्तिमा १२० दिन म्याद भएको यस अस्पतालको नाममा जारी भएको विड वण्ड पेश गर्नुपर्नेछ।
- 3. खरीदभएको बोलपत्र/दरभाउपत्र तपशिलमा उल्लेखित मिति र समयमा सार्वजनिक खरिद अनुगमन कार्यालयको Website http//www.<u>bolpatra.gov.np</u>मा गई बैंक भौचर, विडवन्ड, BOQ समेतका कागजात Scan गरी Upload गरी सक्नुपर्नेछ । Upload भएका बोलपत्र तपशिलमा उल्लेखित मिति र समयमा यस अस्पतालमा बोलपत्रदाता वा निजको प्रतिनिधिको रोहवरमा खोलिने छ । प्रतिनिधि उपस्थित नभएपिन वोलपत्र खोल्न वाधा पर्नेछैन । बोलपत्र खरीद गर्ने वाबोलपत्र खोल्ने अन्तिम दिन विदा पर्न गएमा सोको भोलिपल्ट क्रमश उक्त कार्य हुनेछ ।
- ४. बोलपत्रदाताले कवोल गरेको प्रत्येक आइटमको दररेटलाई (कर बाहेक) अंक र अक्षरमा स्पष्टसँग उल्लेख गर्नुपर्नेछ । अंक र अक्षरमा उल्लेखित दररेटमा भिन्नता भएमा अक्षरलाई मान्यता दिइने छ ।
- ४. सार्वजिनक खरिद अनुगमन कार्यालयको Website http//www.<u>bolpatra.gov.np</u>को माध्यमबाट प्राप्त भएका बोलपत्र/दरभाउपत्र मात्र यस अस्पताललाई मान्यहुनेछ । साथै हार्डकिप टेन्डर खुल्ने मिति देखि ५ दिनभित्र यस अस्पतालको प्रशासन कार्यालय खुल्ने समयमा पेश गर्नुपर्नेछ ।
- ६. संभौता गरी तपशिल अनुसारको मेशिनरी औजार तथा उपकरण सप्लाई नगर्ने फर्मलाई नियमानुसार कालो सूचीमा दर्ताको लागि सम्वन्धित निकायमा लेखि पठाइनेछ ।
- ७. थप जानकारीको लागि यस अस्पतालको खरिद इकाई शाखामा सर्म्पक राख्न सिकनेछ।
- भार्त रहेको,रीत नपुगेको तथाम्याद नाघी आएको वोलपत्रउपर कुनै कारवाही हुनेछैन । एक फर्म वा सस्थाको नाममा खरिद गरिएको वोलपत्र अर्को फर्मको नामवाट जम्मा गर्न पाईने छैन ।
- ९. बोलपत्रदाताले आफू कालो सूचीमा नपरेको भनी स्वयं घोषणा पेश गर्नुपर्नेछ।
- ११. यसमा नपरेका अन्य कुराहरु सार्वजनिक खरीद ऐन २०६३ र सार्वजनिक खरीद नियमावली २०६४ अन्सार हनेछ ।





सि	ठे.नं.	विवरण	धरौटी रकम	दरभाउतथा	टेण्डर दाखिलागर्ने	दरभाउतथा	टेण्डर दस्तुर	कैफिय
. नं				टेण्डर	अन्तिममिति	टेण्डर खोल्ने	दाखिलागर्ने बैक	त
				दस्तुर		मिति	तथाखाता नं.	
٩	४७।०७३।०७४			90001-	३० औंदिनको	३१ औंदिनको	रा.वा.बैक भरतपुर,	
		Mortuary			प्रा०० वजे सम्म	२।०० वजे	च.हि.नं	
		freezer(15 bodies)					१३१००००९६	
			155000.00				<b>७</b> ०१	
			155000.00					





The Statement of Compliance must be substantiated with authenticated catalogue/data sheet/ manual with the page number of original catalogue/datasheet/manual of the relevant parameters indicated.

#### **Technical Specification of Mortuary Freezer (15 bodies)**

SN	Technical Specification	Bidder's Offer	Catalogue Pg. No	
	Mortuary Freezer (15 bodies)			
	Brand:			
	Model:			
	Manufacturer:			
	Country of origin:			
1.	<b>Description of function:</b> Refrigerator freezer that is used to store the corpses for short and long term.			
2.	bodies or more	5		
3.	Technical Specifications			
4.	Temperature: Temperature must be maintained within the range from 2 8 degree Celsius.	-		
5.	The freezer must have a digital microprocessor based temperature control			
6.	<u>Display:</u> The system must have a clear LCD/LED display in the front			
7.	Construction: Chamber Outer Body shall be constructed with M.S with powder coated and Inner with 304 Stainless Steel or bette material.	r		
8.	The system shall have an insulation of high grade poly Uretherane.			
9.	Shall have steel doors having magnetic gasket and external lock outside the door.	I		
10	Shall have trolley mounted on telescopic rail and able to slide in and out.			
11	The freezer compressor shall be marked with ISI mark and the refregirant used must be CFC Free R404a			
12	. Shall have automatic defrosting feature.			
13	Shall have an internal drainage system and integral temp gauge.			





	1	Disease
14.	Shall have air cooled hermatically sealed refrigeration system	
15	Shall have audible and visible alarms for high and low	
15.	temperature and other mechanical errors.	
	Accessories, spares and consumables	
	All standard accessories, consumables and parts required	
	to operate the equipment, including all	
16.	standard tools and cleaning and lubrication materials, to be	
	included in the offer. Bidders must	
	specify the quantity of every item included in their offer (including items not specified above).	
	Power supply: 220-240V/ 50 Hz AC Single phase fitted with	
17.	appropriate plugs. The power cable must be minimum 3	
	metres long.	
	Suitable Automatic Voltage regulator/stabilizer meeting	
	international standards must be	
	international standards must be	
	supplied. Broad specifications are: Automatic Type Input	
	150-280V, Output 220 V +/- 7 %, 50	
	Hz. Single phase, AC with automatic 2-4 sec Cut Off and 6-9	
18.	minutes restart delay. Quick start	
	anno anno anta fambanno aireath a daoine dheach a ceann an ann an ann an ann an ann an ann an a	
	arrangements for bypassing the start delay. Suitable MCB	
	on input voltmeter and indicators on	
	Front Panel. Input Pore Cable with 15 A Plug and six way	
	output terminal strip for two outlets	
	output terriman strip for two outlets	
19.	Terms And Conditions:	
15.		
20.	Must submit ISO 9001 or ISO13485:2003/AC:2007 <b>AND</b>	
	Shall meet IEC 60335-1and -2-24 General requirements of	
21.	electrical safety.	
	electrical safety.	
	Must provide user training (including how to use and	
22.	maintain the equipment).	
	, , ,	
23.	The supplier must submit the original brochure or e-copy.	
	The local agent should be available instantly for the after	
24.	The local agent should be available instantly for the after	





	sales services	
25.	Warranty should be 2 years from the date of installation of the machine and another 2 years of free service should be provided.	
26.	During the warranty period supplier must ensure planned preventive maintenance (PPM) along with corrective / breakdown maintenance whenever required.	
27.	The onsite Operational training should give by qualified technician to the users (how to use) including maintenance training to the hospital's biomedical engineer and technician.	
28.	Hard copy and soft copy of both user manual and service manual in English shall be provided at the time of installation.	



# नेपाल सरकार स्वास्थ्यमन्त्रालय भरतपुर अस्पताल , चितवन

**Bill of Quantity** 

S.No.	Particular	Quantity	Unit	Per Rate		Total Amount (Rs.)	Remarks
				In figures	In words	iotai Amount (NS.)	Remarks
	Mortuary freezer(15 bodies)	1	set				
1		13% Va	t				
		Total Amou	unt				
	Name of Propritor :				Signature :		
	Phone :				Mobile :		
	Address:				Fax :		
	Date :				E-mail :	Stamp	

